

製品情報

見本

製品コード:	4-DP-004 (旧: 10603)
製品名:	ダニ (Dp) 抽出物 (50%グリセロール含有)
抗原:	ヤケヒョウヒダニ
原料:	<i>Dermatophagoides pteronyssinus</i>
学名:	<i>Dermatophagoides pteronyssinus</i>
成分:	ヤケヒョウヒダニ培地粗抽出物 (0.45 µm 濾過済) Phosphate buffered saline (pH7.4) 0.05 % ProClin™ 950 (防腐剤) 含有
状態:	グリセロール50%含有溶液
容量:	2 mL (1 mL/tube x 2本)
濃度:	Der p 1 7.5 µg/mL
保存:	-20°C
注意:	グリセロールを含むため、-25°C以下だと凍結する可能性があります。 本品は研究用試薬です。ヒトへの使用はできません。 本品にはアレルギー性があります。取り扱う際は、白衣、防護眼鏡、マスクを使用した上で、 皮膚および粘膜に付着しないように注意してください。

分析証明書

Lot No. XXXXXXXXXX

1. アレルゲン濃度	規定あり: XXX µg/mL *1 実測値はロットごとに異なります。	測定法: サンドウィッチELISA法 (Der p 1)
2. 総蛋白質濃度	規定なし: -	測定法: BSAを指標にしたLowry法

ITEA株式会社 東京環境アレルギー研究所
東京都文京区白山1-33-18 白山NTビル
Tel: 03-3526-2031 Fax: 03-3526-2032

発行責任者: 所長 XXXXX

品質保証担当者: 品質保証課 XXXXX

*2 製品に添付する際には、社判、責任者印、品質保証担当者印が押印されます。

発行日: XXXX年 XX月 XX日

PRODUCT INFORMATION

SAMPLE

Product code: 4-DP-004 (Former code: 10603)
Product name: House Dust Mite Extract, Whole Culture (Dp) (Contains 50% glycerol)
Antigen: *Dermatophagoides pteronyssinus*
Material: *Dermatophagoides pteronyssinus*
Scientific name: *Dermatophagoides pteronyssinus*
Constituents: Extracts from house dust mite (Dp) culture medium (filtered with 0.45 µm)
Phosphate buffered saline (pH7.4) Containing 0.05 % ProClin™ 950 (preservative).
Appearance: Solution containing 50% glycerol.
Volume: 2 mL (1 mL/tube x 2本)
Concentration: Der p 1 7.5 µg/mL
Storage: -20°C
It may freeze at temperatures below -25°C because of glycerol containing.
Note: This product should be used for research use only. Cannot be used for treatment and diagnosis purposes.
When handling it, wear a lab coat, protective glasses, a mask and gloves.
Be careful not to let it contact with your skin or mucous membranes because of its allergenicity.

CERTIFICATE OF ANALYSIS

Lot No. XXXXXXXXXX

1. Allergen concentration Specified: XXX µg/mL Measurement method: Sandwich ELISA
*1 Actual measured value varies from lot to lot. (Der p 1)
2. Total protein concentration Not specified: -

Institute of Tokyo Environmental Allergy
Hakusan NT Building, 1-33-18 Hakusan, Bunkyo-ku, Tokyo
Tel: +81-3-3526-2031 Fax: +81-3-3526-2032

ITEA Director: XXXXXQuality Assurance: XXXXX

Date of issue: mm dd, yyyy